


## Separation of Ascorbic acid and Nicotinamide

Data# B-045



Column	Mightysil RP-18 GP 250-4.6mmØ 5µm	
Eluent	10mM Disodiumhydrogen phosphate containing 5mM Tetrabutylammonium sulfate (pH 6.8)	
Flow Rate	1.0mL / min.	
Detection	UV254nm	
Temperature	Room Temperature	
Sample Volume	10µl	
Detected peaks	1. Ascorbic acid	0.1 mg / ml
	2. Nicotinamide	0.1 mg / ml