


## Separation of Curcumins

Data# B-095



Column	Mightysil RP-18 GP 250-4.6mmØ 5µm
Eluent	Acetonitrile / Water / Acetic acid = 45 / 55 / 1
Flow Rate	1.0mL / min.
Detection	UV422nm
Temperature	40
Sample Volume	5µl
Detected peaks	Mixture of Bisdemethoxycurcumin Demethoxycurcumin Curcumin