


Mightysil Application DATA

Separation of Biotin

Data# B-108



Column	Mightysil RP-18GP 150-4.6, 5 μ m
Eluent	50mM KH ₂ PO ₄ (pH3.0)/ Acetonitrile= 9/ 1
Flow Rate	1.5mL/min
Detection	UV210nm
Temperature	40°C
Sample Vol.	100 μ L
Sample	Biotin (2.2 μ g/mL)



Kanto Kagaku