


Separation of 3 Agricultural Chemicals

Data# E-036



Column	Mightysil RP-18 GP 150-4.6mmØ 5µm	
Eluent	A; 10mM KH ₂ PO ₄ (Asusted pH3.50 by addition of H ₃ PO ₄) B; Acetonitrile	
Flow Rate	1.0mL / min.	
Detection	UV254nm 0.02 AUFS	
Temperature	40°C	
Sample Volume	10µl	
Detected peaks	1. Destylpyrazolate	2.5 µg / ml
	2. Dithianone	2.5 µg / ml
	3. Triflumizole	2.5 µg/ml