


Separation of Quassins

Data# F-064



Column	Mightysil RP-18 GP 150-4.6mmØ 5µm	
Eluent	Acetonitrile / Water = 35 / 75	
Flow rate	1.0 ml / min.	
Detection	UV254nm	
Temperature	30°C	
Sample Volume	10µl	
Detected peaks	1.2. : Quassin	100µg/ml
	3. p-Methylacetophenon	12.5µg/ml