


## Separation of Presevatives -4

Data# F-075



Column	Mightysil RP-18 GP 150-4.6mmØ 5µm	
Eluent	5mM Citric acid buffer*/ Acetonitrile / Methanol = 70/20/10 *: Consist of 0.7g/l Citric acid monohydrate and 0.6g/l Trisodium citrate dehydrate	
Flow rate	1.0 ml / min.	
Detection	UV230nm	
Temperature	40°C	
Sample Volume	10µl	
Detected peaks	1. Benzoic acid	100 µg/ml
	2. Sorbic acid	100 µg/ml
	3. Salicylic acid	100 µg/ml
	4. Dehydroacetic acid	100 µg/ml