



Mightysil Application DATA

Separation of Benzoyl Peroxide in Flour


Standard 0.5 μ g / mL



Flour



Flour + Benzoyl Peroxide 0.5 μ g / mL



Column	Mightysil RP-18GP 250-4.6mm \varnothing 5 μ m
Eluent	Acetonitrile / Water = 55 / 45
Flow Rate	1.0mL/min.
Detection	UV235nm
Temperature	40 $^{\circ}$ C
Sample Vol.	20 μ L
Preparation	The Acetonitrile is added to flour , and filtered



Kanto Kagaku